



- Meet DIN878
- With limit pointers
- Optional accessory: backs (page 129~130), contact points (page 126~128)



2886-3F

COMPACT DIAL INDICATOR

Code	Range	Graduation	Accuracy	Hysteresis	Range/rev	Dial reading	Remark
2886-3F*	3mm	0.01mm	12 μ m	3 μ m	0.5mm	0-50	flat back

* Supplied with manufacturer inspection certificate traceable to NIST USA

