

BACK PLUNGER TYPE DIAL INDICATOR



handle (included)



- Supplied with handle
- With limit pointers
- Optional accessory:
contact points (page 126~128)



2357-5

Code	Range	Graduation	Accuracy	Hysteresis	Range/rev	Dial reading
2357-5 *	5mm	0.01mm	15µm	5µm	1mm	0-100

* Supplied with manufacturer inspection certificate traceable to NIST USA

